

磁振影像學MRI 鬆弛時間



盧家鋒 教授

國立陽明交通大學
生物醫學影像暨放射科學系
alvin4016@nycu.edu.tw



Procedure of MRI

- Alignment (magnetization) B_0
- Precession $\omega_0 = \gamma B_0$
- Resonance (given B_1 by RF with ω_2) $\omega_1 = \gamma B_1$, $B_1 \perp B_0$
 - The most effective resonance is produced when $\omega_0 = \omega_2$
- MR signal (EMF, relaxation time)
- Imaging (Pulse sequencing)
 - Tissue Contrast: Image weighting
 - Spatial localization: Spatial Encoding



<http://cflu.lab.nycu.edu.tw>. Textbook: MRI The Basics, Hashemi et al.

2025/9/14


2



本週課程內容 <http://cflu.lab.nycu.edu.tw>

- 鬆弛時間
- 組織鬆弛時間

- MRI The Basics (3rd edition)
 - Chapter 4: T1, T2, and T2*
- MRI in Practice, (4th edition)
 - Chapter 1: Basic Principles
 - Chapter 2: Image Weighting and Contrast



<http://cflu.lab.nycu.edu.tw>. Textbook: MRI The Basics, Hashemi et al.

2025/9/14

3

鬆弛時間T1, T2, T2*

Relaxation Time

<http://cflu.lab.nycu.edu.tw>. Textbook: MRI The Basics, Hashemi et al.

2025/9/14

4



Relaxation

- During relaxation, hydrogen nuclei give up absorbed RF energy and the NMR returns to B_0 direction.
- The magnetic moments of hydrogen lose coherency due to dephasing.
- 1. Recovery of longitudinal magnetization (M_z) \rightarrow T1 recovery
- 2. Decay of transverse magnetization (M_{xy}) \rightarrow T2 decay

<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.


2025/9/14

5



Relaxation


T1 recovery
(slower)



fat regains longitudinal magnetization rapidly

time

T2 decay
(faster)



<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14

6



T1, T2, T2*

- Inherent properties: T1 and T2
 - Fixed for a specific tissue at a given B_0 strength
- T2*
 - The effects of T2 and inhomogeneities in the B_0
 - Fixed for a specific tissue within a given external magnetic environment

<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14

7



T1 Relaxation Time

- Relaxation: the spins are relaxing back into their lowest energy state or back to the equilibrium state.
- T1: the longitudinal relaxation time
 - It refers to the time it takes for the spins to realign along the longitudinal (z) axis.
- T1: the spin-lattice relaxation time
 - It refers to the time it takes for the spins to give the energy they obtained from the RF pulse back to the surrounding lattice in order to go back to their equilibrium state.

<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14

8



T1 Characteristics

- Determined by how well the protons can give off their energy to the surrounding lattice.
- The most efficient energy transfer occurs when the **natural motional frequencies (ω)** of the protons are at the **Larmor frequency (ω_0)**.
- If the efficiency of energy transfer from the protons to the lattice is increased, the T1 becomes shorter.
- Range of $\omega(\text{H}_2\text{O}) \gg \omega_0$, $\omega(\text{solids}) < \omega_0$, $\omega(\text{fat}) \approx \omega_0$

<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14

9



Longer T₁ at Higher Field B₀

~~Over range of field strengths used for clinical MRI (0.3 T – 3.0 T)~~

- A higher magnetic field is stronger, it will pull the net magnetization vector back to its original position in the z-direction more quickly, producing shorter T1.
- **What is the number of resonant protons that are available to transfer energy to the "lattice"?**
- Increase the magnetic field strength B₀ → the Larmor frequency increases → fewer protons experience an oscillating magnetic field at or near the Larmor frequency → fewer protons available to transfer energy efficiently to the lattice, and the T1 time is lengthened.


<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14

10


Distribution of Molecular Motion in Tissue

<https://mriquestions.com/bo-effect-on-t1-t2.html>



Field Strength Dependence of T1


<https://mriquestions.com/bo-effect-on-t1-t2.html>





T1 Recovery

- After the RF pulse is turned off ...
 - The M_{xy} component of the magnetization vector decreases rapidly.
 - The M_z component slowly recovers along the z axis.
- $M_z(t) = M_0(1 - e^{-t/T_1})$



<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14


13



T2 Relaxation Time

- T_2 : transverse relaxation time
- T_2 : spin-spin relaxation time
- T_2 decay depends only on
 - Spin-spin interactions
- T_2^* decay depends on both
 - External magnetic field
 - Spin-spin interactions

$T_2 > T_2^*$, T_2^* decay is quicker than T_2 decay



<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14

14



T2 Characteristics

- Determined by how fast the proton spins in that tissue dephase.
- If they dephase rapidly, we get a short T2.
- If they dephase slowly, we get a longer T2.

<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.


2025/9/14

15



Dephasing

- ΔB_{int} • Interactions between individual spins (internal inhomogeneities)
 - When two spins are next to each other (one is aligned with B_0 and the other is against it)...
 - $\omega(\text{proton } \#1) = \gamma(B_0 + \Delta B)$
 - $\omega(\text{proton } \#2) = \gamma(B_0 - \Delta B)$
 - Depends to a degree on *the proximity of the spins* to each other.
- ΔB_{ext} • External magnetic field inhomogeneity
 - Protons in different locations precess at different frequencies.



<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.


2025/9/14

16



T2 Relaxation Time

- As the M_z recovers, the transverse vector M_{xy} decays at a rate characterized by T_2
- $M_{xy}(t) = M_0 e^{-t/T_2}$
- T_2 decay occurs 5 to 10 times more rapidly than T_1 recovery.



<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.


2025/9/14

17



T1, T2, T2*

- $T_1 > T_2 > T_2^*$
- $1/T_2^* = 1/T_2 + \gamma \Delta B_{ext}$



<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.


2025/9/14

18



The Received Signal

- A oscillated magnetic field causes movement of electrons, i.e., the current (signal).
- A single RF coil can only detect the component of magnetization along a specific axis on the XY plane.



<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.


2025/9/14

19



Free Induction Decay

- The oscillating, decaying signal is called a free induction decay (FID).
- After we turn off the RF pulse...
 - The spins begin to precess *freely*.
 - The spins *induce* a current in the receiver coil.
 - The signal starts to *decay* with time.
- $M_{xy}(t) = M_0 e^{-t/T_2^*} (\cos \omega_0 t)$



<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14

20

組織鬆弛時間

Tissue Relaxation Time

<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.


2025/9/14

21



T1 and T2

- T1 is defined as the time when 63% of the longitudinal magnetization has recovered (with $3*T1=95\%$ recovery).
- T2 is defined as the time when 63% of the transverse magnetization has decayed (with $3*T2=95\%$ decay).




<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14

22

T1 and T2 Characteristics


- T1 (longer T1, slower M_z recovery)
 - $H_2O > Solid\ tissue > Fat$
- T2 (longer T2, slower M_{xy} decay)
 - $H_2O > Fat > Solid\ tissue$



<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.


2025/9/14

23



T2 Water vs. Fat

Water





<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14


24

Fat





Water




Although the hydration layer water has much shorter T1 than bulk water, it also has shorter T2 (fast signal decay).

<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14

25



T1 and T2 Characteristics: Brain Tissues

- T1 (longer T1, slower M_z recovery)
 - Cerebrospinal fluid (CSF) > Gray matter (GM) > White matter (WM)
- T2 (longer T2, slower M_{xy} decay)
 - CSF > GM > WM

Table 6-1 T1, T2, and Proton Density of Brain Tissues at 1.5 T*

| | T1 (msec) | T2 (msec) | N(H) |
|--------------|-----------|-----------|------|
| White matter | 510 | 67 | 0.61 |
| Gray matter | 760 | 77 | 0.69 |
| Edema | 900 | 126 | 0.86 |
| CSF | 2650 | 180 | 1.00 |

* Stark and Bradley, p. 44.

<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.

2025/9/14

26



Video Demonstration magritek-youtube channel

Video 4 3:23 FID & Larmor frequency




Video 5 2:50/3:43 B0 inhomogeneity → T2*



<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.


2025/9/14

27




Video Demonstration magritek-youtube channel


Video 5 4:30 shimming → T2



without shimming → T2*



with shimming → T2



time frequency

<http://cflu.lab.nycu.edu.tw>, Textbook: MRI The Basics, Hashemi et al.



2025/9/14

28



Procedure of MRI

- Alignment (magnetization) B_0
- Precession $\omega_0 = \gamma B_0$
- Resonance (given B_1 by RF with ω_2) $\omega_1 = \gamma B_1$, $B_1 \perp B_0$
 - The most effective resonance is produced when $\omega_0 = \omega_2$
- MR signal (EMF, relaxation time)
- Imaging (Pulse sequencing)
 - Tissue Contrast: Image weighting
 - Spatial localization: Spatial Encoding



THE END

alvin4016@nycu.edu.tw